



China recognizes international
mutual recognition testing
TESTING CANS L1184

TEST REPORT

NO: 23-000466

Sample Name: KN95 Mask

Applicant: Qingdao Yunxian Protective Equipment Co., Ltd.

Manufacturer: /

Test Category: Commissioning Test



Qingdao product Quality Testing Research Institute
National Ecological Textiles Quality Inspection and Testing Center

Qingdao product Quality Testing Research Institute
National Ecological Textiles Quality Inspection and Testing Center



TEST REPORT



Report No.:23-000466
Verification Code: DBF204194F

2023011212221505000466

Page 1 of 3

Sample Description*	KN95 Mask		Test Type*	Commissioning Test	
Specification*	/		Trade mark*	/	
Applicant*	Qingdao Yunxian Protective Equipment Co., Ltd.		Phone number*	13675328745	
Name/ * Address of Manufacturer	/		Phone number*	/	
Entrusted Person	王凯		Sample Grade*	KN95	
Sampling Location	/	Sampling Base		Sampling Quantity	/
Sample Status	White		Sample Number	2023011212221505000466	
Style/Article No.*	/				
Quantity of * entrusted samples	40 PCS		Date of Sample receipt	2023-01-12	
Test Date	2023-01-12 — 2023-02-15				
Test Items	Filtration efficiency (without pretreatment), inhalation resistance (without pretreatment), exhalation resistance (without pretreatment), headband (without pretreatment), appearance inspection (without pretreatment)				
Inspection basis /Judgment principle	GB 2626-2019 《Respiratory protection - Self-priming filter-type particulate respirator》				
Test Results	According to the client's requirements, the test is carried out according to the above test basis. Please see the test result summary page for details. (Special stamp for inspection report) Issue date: February 15, 2023				
Note	The contents marked with an asterisk (*) are information provided by the client, and the inspection results are only responsible for the samples provided.				
Approver: Reviewer: Preparer/Editor-in-Chief:					



TEST REPORT


Report No.:23-000466

2023011212221505000466

Page 2 of 3

Summary of Test Results					
Serial	Test Items name		Technical requirements	Test Results	Evaluation indicators
1	Filtration efficiency (without pretreatment)	%	Unprocessed: Saline Medium ≥ 95.0	1# 99.8	Qualified
			Unprocessed: Saline Medium ≥ 95.0	2# 99.8	
			Unprocessed: Saline Medium ≥ 95.0	3# 99.7	
			Unprocessed: Saline Medium ≥ 95.0	4# 99.7	
			Unprocessed: Saline Medium ≥ 95.0	5# 99.5	
			Unprocessed: Saline Medium ≥ 95.0	6# 99.6	
			Unprocessed: Saline Medium ≥ 95.0	7# 99.8	
			Unprocessed: Saline Medium ≥ 95.0	8# 99.7	
			Unprocessed: Saline Medium ≥ 95.0	9# 99.8	
			Unprocessed: Saline Medium ≥ 95.0	10# 99.7	
			Unprocessed: Saline Medium ≥ 95.0	11# 99.4	
			Unprocessed: Saline Medium ≥ 95.0	12# 99.6	
			Unprocessed: Saline Medium ≥ 95.0	13# 99.6	
			Unprocessed: Saline Medium ≥ 95.0	14# 99.4	
			Unprocessed: Saline Medium ≥ 95.0	15# 99.5	
2	Inspiratory resistance (without pretreatment)	Pa	Unprocessed: Disposable Face Mask, No Exhalation Valve ≤ 210	16# 181.1	Qualified
			Unprocessed: Disposable Face Mask, No Exhalation Valve ≤ 210	17# 192.4	
3	Inspiratory resistance (without pretreatment)	Pa	Unprocessed: Disposable Face Mask, No Exhalation Valve ≤ 210	18# 90.3	Qualified
			Unprocessed: Disposable Face Mask, No Exhalation Valve ≤ 210	19# 102.4	
4	Appearance inspection (without pretreatment)	/	Without pretreatment, 10N for 10s, no slippage or breakage should occur.	10N for 10s without slipping or breaking	Qualified
5	Appearance inspection (without pretreatment)	/	The surface of the untreated sample should not be damaged, deformed or have other obvious defects; The labeling and various information provided by the manufacturer shall comply with Article 7 of the standard.	There is no damage, deformation or other obvious defects on the surface of the sample.	/
Test Description	The above inspection items are conducted in accordance with the following standard: GB 2626-2019 Respiratory protection—Self-inhalation filter-type particulate respirators				
Note	/				





Report No.:23-000466

Page 3 of 3

2023011212221505000466



Notes

- 1、报告无“检验检测专用章”无效。
- 2、检验报告未经本院/中心书面批准不得复制（全文复制除外），经特许复制的报告需重新加盖“检验检测专用章”，否则无效。
- 3、报告无主检、审核、批准人签章无效。
- 4、报告涂改无效。
- 5、考虑到您的利益，如果对委托检验报告结果有异议，请在收到报告之日起十五日内向检验单位提出。
- 6、送样委托检验，检验结果仅对来样负责。未经本院/中心同意委托人不得擅自使用检验结果进行不当宣传。

- 1、The report is invalid without Inspection report special stamp.
- 2、Do not copy the report unless specially permitted. The copy must be sealed with inspection report special stamp, otherwise it would be invalid.
- 3、The report would be invalid if there is no signature of the major inspector, verifier or authorizer.
- 4、The report would be invalid if altered.
- 5、Any question with the Commissioned inspection report should be submitted to the inspection unit within 15 days from receiving the report. After the specified date any request would be refused.
- 6、Sample offered inspection results refer only to the sample(s) tested. Do not use the results for improper propaganda.

地址：青岛市崂山区深圳路173号
邮编：266061
电话：0532-83871780
传真：0532-88918100
网址：www.qtc.org.cn
E-mail：qdzjs@qtc.org.cn

Add: No.173 Shenzhen Road Qingdao China
P.C.: 266061
Tel: 0532-83871780
Fax: 0532-88918100
Website: www.qtc.org.cn
E-mail: qdzjs@qtc.org.cn

